

510(k) Summary for the Bridge SE Biliary Self-Expanding Stent Delivery System

JAN 14 2002

**510(k)
Summary**

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of 21 C.F.R. § 807.92.

Submitter

Medtronic AVE, Inc., Peripheral Technologies
2170-A Northpoint Parkway
Santa Rosa, California 95407

Contact Person

John Riolo
Vice President, Regulatory Affairs and Quality Assurance
Phone: (707) 541-3271
FAX: (707) 566-1159
e-mail: john.riolo@medtronic.com

Date Prepared

December 20, 2001

Trade Name

Bridge SE Biliary Self-Expanding Stent Delivery System ("Bridge SE")

**Common
Name**

Biliary Stent and Delivery System

**Classification
Name**

Biliary Catheter and Accessories

**Device
Classification**

Classification: Class II
Classification Panel: 78FGE
Regulation Number: 21 C.F.R. §876.5010

**Predicate
Device**

Bridge SE Biliary Self-Expanding Stent Delivery System
K011080; cleared 10/11/2001

**Performance
Standards**

Performance standards have not been established by the FDA under section 514 of the Federal, Food, Drug and Cosmetic Act

510(k) Summary for the Bridge SE Biliary Self-Expanding Stent Delivery System

Device Description	The subject device is a system consisting of a single-lumen, tubular, self-expanding nickel-titanium ("nitinol") stent pre-loaded onto a sheathed delivery catheter. The delivery system is designed to deliver the stent to the stricture site via transhepatic access. Once positioned at the stricture site, the sheath is withdrawn and the stent is released. Upon release, the stent expands and conforms to the inner lumen of the biliary duct. The device has two radiopaque markers to aid in stent placement during fluoroscopy.
Indications for Use	The Bridge SE Biliary Self-Expanding Stent Delivery System is indicated for use in the palliation of malignant neoplasm in the biliary tree.
Technological Characteristics	The Bridge SE Biliary Self-Expanding Stent Delivery System ("Bridge SE") is substantially equivalent to the currently marketed Bridge SE Biliary Self-Expanding Stent Delivery System. The subject and predicate stents are technologically similar and are intended for palliation of malignant neoplasms in the biliary tree. The subject and predicate stents are constructed of biocompatible materials. The subject and predicate stents are delivered percutaneously via sheathed delivery systems. The subject device offers additional stent sizes and a longer catheter length. The subject and predicate stents are comparable and are intended to meet clinical needs. The difference between the subject and predicate devices are minor and are not relevant to the ability of the subject device to palliate malignant neoplasms in the biliary tree.
Nonclinical Performance	Preclinical testing was conducted to confirm the safe and effective performance of the Bridge SE Biliary Self-Expanding Stent Delivery System as well as the biocompatibility of the device.
Sterilization	The Bridge SE Biliary Self-Expanding Stent Delivery System is provided sterile. The device is not intended for reuse or resterilization.
Conclusion	The Bridge SE Biliary Self-Expanding Stent Delivery System is substantially equivalent to the currently cleared and marketed device and meets the clinical needs of the physicians.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JAN 14 2002

Mr. Kevin Drisko
Regulatory Affairs Manager
Medtronic AVE, Inc.
Medtronic Peripheral Technologies
2170-A Northpoint Parkway
Santa Rosa, California 95407

Re: K014205

Trade/Device Name: Medtronic AVE Bridge SE Biliary Self-Expanding Stent Delivery System
Regulation Number: 21 CFR §876.5010
Regulation Name: Biliary catheter and accessories
Regulatory Class: II
Product Code: 78 FGE
Dated: December 20, 2001
Received: December 21, 2001

Dear Mr. Drisko:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act and the limitations described below. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

The Office of Device Evaluation has determined that there is a reasonable likelihood that this device will be used for an intended use not identified in the proposed labeling and that such use could cause harm. Therefore, in accordance with Section 513(i)(1)(E) of the Act, the following limitation must appear in the Warnings section of the device's labeling:

The safety and effectiveness of this device for use in the vascular system
have not been established.

Furthermore, the indication for biliary use must be prominently displayed in all labeling, including pouch, box, and carton labels, instructions for use, and other promotional materials, in close proximity to the trade name, of a similar point size, and in bold print.

Page 2 – Mr. Kevin Drisko

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and permits your device to proceed to the market. This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification if the limitation statement above is added to your labeling, as described.

Please note that the above labeling limitations are required by Section 513(i)(1)(E) of the Act. Therefore, a new 510(k) is required before these limitations are modified in any way or removed from the device's labeling.

If you desire specific information about the application of other labeling requirements to your device (21 CFR Part 801 and additionally Part 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4616. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Bernard E. Statland", is written over the typed name.

Bernard E. Statland, M.D., Ph.D.

Director

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure


510(k) Number (if known): K014205

Device Name: Medtronic AVE Bridge SE Biliary Self-Expanding Stent Delivery System

FDA's Statement of the Indications For Use for device:

The Medtronic AVE Bridge SE Biliary Self-Expanding Stent Delivery System is indicated for use in the palliation of malignant neoplasms in the biliary tree.

Prescription Use ☒ OR Over-The-Counter Use ☐
(Per 21 CFR 801.109)


(Division Sign-Off)
Division of Reproductive, Abdominal,
and Radiological Devices
510(k) Number K014205